



CE Declaration of Conformity

We, CellPoint Scientific Inc., declare under our sole responsibility that the following product:

Germinator 500 Dry Sterilizer
Model: 5-1460-EU, 5-1460-UK and 5-1460-IT

Has been tested and found in conformity with the following harmonized European standards:

Medical Devices EN 60601-1-2 3rd edition

Therefore comply with the essential requirements and provisions of the following Directives of the European Parliament and European Council:

Electromagnetic Compatibility EMC-Directive: 2014/30/EC
Electrical Equipment Low Voltage Directive: 2014/35/EC

These conclusions are based on the following test report:

991004/GERM/EMC and 041800/AC187/LV



CE-Test, Qualified Testing bv
Klotoweg 363
3047 BG Rotterdam
The Netherlands

EU Authorized Representative: Eisenhut Instrumente GmbH
Schildeckstr 28-30
78665 Frittlingen
Germany

Raymond A Hodgson
General Manager
June 21, 2016